

OSCE Checklist: Recording an ECG

| | Gather appropriate equipment | |
|-----|--|--|
| 1 | Gather appropriate equipment | |
| | | |
| 2 | Wash your hands and don PPE if appropriate | |
| 3 | Introduce yourself to the patient including your name and role | |
| 4 | Confirm the patient's name and date of birth | |
| 5 | Briefly explain what the procedure will involve using patient-friendly language | |
| 6 | If the patient is a female, explain the need for a chaperone | |
| 7 | Gain consent to proceed with the ECG recording | |
| | Adequately expose the patient's chest for the procedure (offer a blanket to allow exposure only when required). Exposure of the patient's lower legs and wrists is also necessary to apply the limb leads. | |
| | Ask the patient to lay on the clinical examination couch with the head of the couch at a 45° angle | |
| 10 | Ask the patient if they have any pain before proceeding with the clinical procedure | |
| Ch | est electrode and lead placement | |
| 11 | Apply V1 at the 4th intercostal space at the right sternal edge | |
| 12 | Apply V2 at the 4th intercostal space at the left sternal edge | |
| 13 | Apply V3 midway between the V2 and V4 electrodes | |
| 14 | Apply V4 at the 5th intercostal space in the midclavicular line | |
| 15 | Apply V5 at the left anterior axillary line at the same horizontal level as V4 | |
| 16 | Apply V6 at the left mid-axillary line at the same horizontal level as V4 and V5 | |
| Lin | nb electrode and lead placement | |
| 17 | Apply the red lead on the right forearm, proximal to the wrist | |
| 18 | Apply the yellow lead on the left forearm, proximal to the wrist | |
| 19 | Apply the green lead on the left lower leg, proximal to the ankle | |
| 20 | Apply the black lead on the right lower leg, proximal to the ankle | |
| Re | cording the trace | |
| 21 | Turn on the ECG machine and ensure ECG paper has been loaded into the machine | |
| 22 | Double-check all the electrodes are attached in the appropriate locations | |
| | Politely ask the patient to remain still and not talk during the recording as muscle activity can cause an artefact which obscures the ECG trace of myocardial activity | |
| | Press the appropriate button on the ECG machine to record the ECG trace. If the ECG trace is poor, double-check the connections to ensure there is good skin contact. | |

| Тс | To complete the procedure | | |
|----|--|--|--|
| 25 | Once an ECG trace has been obtained, switch off the ECG machine | | |
| 26 | Detach the ECG leads from the electrodes and then remove the electrodes carefully, warning the patient this may feel uncomfortable | | |
| 27 | Explain to the patient that the procedure is now finished | | |
| 28 | Thank the patient for their time | | |
| 29 | Dispose of PPE appropriately and wash your hands | | |
| 30 | Label the ECG with the patient's details | | |
| 31 | Document your findings in the patient's notes | | |

Read the full guide at geekymedics.com



Download our clinical skills app